


# Medical Devices and Environmental Monitoring



Authored by:

**Jason Kelly, Applications Director**

2025 Refresh



## Overview

Medical devices are any instrument, apparatus, machine, implant, or other similar article intended for use in the diagnosis, treatment, or prevention of disease or other medical conditions. Medical devices can range from simple devices such as thermometers and blood pressure monitors to complex devices such as pacemakers, artificial joints, and diagnostic imaging equipment.

Medical devices are regulated by government agencies such as the U.S.

Food and Drug Administration (FDA) in the United States and the European Medicines Agency (EMA) in Europe to ensure that they are safe and effective for their intended use. The regulatory process involves rigorous testing and evaluation to ensure that the devices meet specific performance standards and do not pose a risk to patients.

# Medical Device

## Classes

The FDA classifies medical devices into three classes based on their level of risk: Class I, Class II, and Class III. Class I devices are low-risk devices such as tongue depressors and bandages, while Class III devices are high-risk devices such as implantable pacemakers and artificial hearts.

In Europe, medical devices are regulated by the European Medicines Agency (EMA) under the Medical Device Regulation (MDR). The MDR also classifies medical devices into classes based on their level of risk: Class I, Class IIa, Class IIb, and Class III.

In the MDR, Class IIa medical devices are those that are considered to have a moderate risk to patients, while Class IIb medical devices have a higher level of risk. The classification is based on a variety of factors, including the duration of use, invasiveness, and the nature of the medical condition being treated.


Examples of Class IIa medical devices include contact lenses, ultrasound imaging systems, and hearing aids, while Class IIb devices include, surgical lasers and infusion pumps.

### MDR Class III Medical Devices:

These are high-risk medical devices that are intended to sustain or support life, or are implanted in the body. Examples of Class III devices include heart valves, pacemakers, and implantable infusion pumps. Class III devices require premarket approval by the FDA, which involves demonstrating safety and effectiveness through clinical studies and other evidence.

Class III medical devices are those that are invasive and are used for a long-term period, such as implantable devices that are intended to remain in the body for more than 30 days. Examples of Class III devices include implantable cardiac pacemakers, artificial heart valves, and implantable neurostimulators.

In contrast, Class IIb medical devices are typically non-invasive or minimally invasive, such as infusion pumps or ultrasound equipment, and are designed to be used for a shorter duration than Class III devices. They still pose a significant level of risk to patients and require a higher level of regulatory scrutiny than Class IIa devices.



Examples of Class I medical devices.

## Class I


These types as low-risk devices. For example, bandages, band aids, wheelchairs, surgical tools and other types such as oxygen masks, hospital beds, tongue depressors.



Examples of Class 2 medical devices.

## Class 2

These include, for example, absorbable sutures, blood transfusion kits, catheters, CT (computed tomography) scanners, blood pressure cuffs, pregnancy test kits, and infusion pumps for intravenous drugs. They are intermediate-risk devices.



Examples of Class 3 medical devices.

## Class 3

Heart pacemakers, defibrillators, cochlear implants, breast implants, high-frequency ventilators, and implanted prosthetics are some examples of Class III devices. They are high risk as they are implanted into the human body or perform critical lifesaving functions.

# Medical Device

## Manufacturing

Medical devices are complex and intricate instruments that require a high degree of precision and quality control in their manufacturing process. The manufacture of medical devices involves a series of steps, starting from the design and development phase to final inspection and packaging. In this article, we will provide an overview of the manufacturing process for medical devices.

### General overview of the manufacturing process

#### Design & Development

The design and development phase is the first step in the manufacture of medical devices. During this phase, engineers and designers work closely to develop the product concept, create 3D models and prototypes, and refine the design until it meets the desired specifications. This process may involve extensive testing and simulation to ensure that the product meets safety and regulatory requirements.

#### Material Selection

Once the design has been finalized, the next step is material selection. Medical devices are made from a wide range of materials, including metals, plastics, ceramics, and composites. The choice of material depends on factors such as the device's intended use, cost, and regulatory requirements. The manufacturer must ensure that the selected materials are safe, biocompatible, and meet the necessary performance specifications.

#### Manufacturing

The manufacturing process for medical devices can vary depending on the type of device being produced. Generally, the manufacturing process includes the following steps:

#### Fabrication

The fabrication process involves the shaping and forming of the device components using various techniques such as machining, molding, casting, or extrusion.

#### Assembly

Once the individual components have been fabricated, they are assembled using various techniques such as welding, bonding, or fastening. The assembly process may be manual or automated, depending on the device's complexity.

#### Testing & Quality

Testing and quality control are critical steps in the manufacturing process. The manufacturer must conduct a series of tests to ensure that the device meets the required specifications and regulatory requirements. These tests may include mechanical, electrical, and biocompatibility testing.

#### Sterilization

Many medical devices require sterilization before they can be used. The manufacturer may use various sterilization techniques such as gamma radiation, ethylene oxide, or steam sterilization, depending on the device's material and intended use.

#### Final Inspection

After the device has been manufactured and sterilized, it undergoes final inspection to ensure that it meets the quality standards and specifications. The device is then packaged and labeled, ready for distribution.

# Particle Monitoring

## in a Medical Device Cleanroom

Unlike some sterile injectable pharmaceutical products where terminal sterilization cannot be performed (as the process would change the properties of the medication and render it unusable) the majority of Class III medical devices undergo terminal sterilization. These techniques remove viable contamination from the product. However there are still potential hazards that still require rigorous monitoring.


### Cleanroom Monitoring:

For Class III device manufacturing the general assembly environment needs monitoring. As well as meeting ISO 14644-1 and 2 requirements a real time monitoring system is best used to monitor these types of cleanrooms. Regular air sampling is also needed. A Risk Assessment would be a good way to decide any other critical areas that need to be directly monitored.

Portable Instruments for particle counter viable particles can also be used but they would require more manual activity compared to the automated real time monitoring system.

### Real Time Monitoring Systems

Consist of multiple environmental sensors connected in zones within the cleanroom deemed to be critical or locations where meaningful data can be recorded. As mentioned previously a formal Risk Assessment is the best starting point to understand where the sensors are actually required. The viable and non



Example of Real Time Monitoring Systems. LWS provides a total solution from one supplier.

viable monitoring points are of high importance and secondary TRH sensors should be placed in places where they are not influenced by local laboratory or manufacturing equipment rather than to get an average TRH reading of the room. For pressure sensors since room pressure is uniform they can be placed on ceiling or walls using sample ports or the sensors can be directly wall mounted if a local pressure visualization is required to enter the room.

The main objective of these systems is to ensure that the environmental conditions are within design specifications for the processes that are conducted in these rooms or locations or separative devices.

FDA requires this data to be recorded and validated and product batch releases depend on this data.

# Medical Device

## Product Monitoring

Certain implantable devices are required to be rinsed using sterile water to clean the parts and then batch samples are taken of the rinsed liquid to test for particulates. Regulatory standards are followed based on acceptable allowable particulates per ml. The reason behind this is that some products such as Class II devices like needles, catheters and any devices that puncture the skin and body need to be checked for large particles. These particles can cause illness or can be even fatal in certain circumstances. Liquid particle counters are used to sample for the presence of particulates on such products.



Liquid Particle Counters (LPCs) are used to monitor the cleanliness of medical device products after parts rinsing. LPCs are specialized instruments that are capable


of measuring the size and number of particles in liquids, such as the rinse water used in medical device manufacturing.

After the parts have been rinsed and dried, a sample of the rinse water may be collected and analyzed using an LPC. The LPC measures the number and size distribution of particles present in the sample, providing information on the cleanliness of the parts and the effectiveness of the rinsing process.

LPCs are often used in cleanroom

environments where the presence of even small particles can pose a risk to patient safety. By monitoring the cleanliness of medical device products using LPCs, manufacturers can ensure that their products meet the necessary quality and safety standards before they are shipped to customers.

It is important to note that LPCs are just one tool that can be used to monitor the cleanliness of medical device products after parts rinsing. Other methods may include visual inspection, microbial testing, or other analytical techniques, depending on the type of device being manufactured and the specific quality control requirements.



*Implantable knee joints*

## Here is an example of an SOP for parts rinsing and testing using a Liquid Particle Counter:

<b>Purpose</b>	The purpose of this SOP is to outline the procedure for rinsing and testing medical device parts using a Liquid Particle Counter (LPC) to ensure that they are clean and free from contaminants.
<b>Scope</b>	This SOP applies to all medical device parts that require rinsing and testing using an LPC.
<b>Equipment &amp; Material</b>	LPC, Rinse solution (e.g. deionized water), Parts to be tested.

### Procedure

<b>Pre-Cleaning</b>	Before parts can be rinsed, they should be pre-cleaned to remove any visible debris or contaminants. This may involve using solvents, detergents, or mechanical cleaning methods.
<b>Rinsing</b>	The parts are then rinsed with a suitable rinse solution to remove any remaining debris or contaminants. Rinse solution should be deionized water or a specific cleaning solution, depending on the type of device being manufactured.
<b>Rinse Water Collection</b>	After rinsing, a sample of the rinse water should be collected for testing using the LPC. The sample should be taken from the rinse water that has passed through the parts to be tested.
<b>LPC Calibration</b>	Prior to use, the LPC should be calibrated according to the manufacturer's instructions to ensure accurate measurement of particle size and count.
<b>LPC Testing</b>	The sample of rinse water should be analyzed using the LPC. The LPC should be operated according to the manufacturer's instructions, and particle count and size distribution data should be recorded.
<b>Result Analysis</b>	The particle count and size distribution data obtained from the LPC should be analyzed to determine the cleanliness of the parts being tested. Acceptable particle count and size distribution limits should be established based on product-specific requirements and industry standards.
<b>Record Keeping</b>	All test results should be recorded and kept on file for reference. The results should include information such as the date and time of the test, the LPC calibration data, the particle count and size distribution data, and any notes or observations.
<b>Results Evaluation</b>	The results of the LPC testing should be evaluated to determine whether the parts being tested meet the necessary quality and safety requirements. If the results are outside of acceptable limits, corrective action should be taken to improve the rinsing process or other aspects of the manufacturing process as necessary.
<b>Conclusion</b>	This SOP provides a standardized procedure for parts rinsing and testing using an LPC. By following this procedure, medical device manufacturers can ensure that their products are clean and free from contaminants, meeting the necessary quality and safety requirements.



# Standards

## for Medical Devices

Medical device manufacturing is subject to various international standards and regulations to ensure the safety and efficacy of medical devices. Here are some of the key standards used in medical device manufacturing:

<b>ISO 13485</b>	This standard specifies requirements for a quality management system for medical device manufacturers. It covers all aspects of the manufacturing process, from design and development to production, installation, and servicing.
<b>FDA Regulations</b>	In the United States, the Food and Drug Administration (FDA) regulates medical devices under the Federal Food, Drug, and Cosmetic Act. Medical devices must meet certain requirements for safety and effectiveness before they can be marketed in the U.S.
<b>EU MDR</b>	The European Union's Medical Device Regulation (MDR) requires manufacturers to meet specific requirements for safety and performance before they can sell medical devices in the EU.
<b>IEC 60601</b>	This standard outlines general safety requirements for medical electrical equipment. It covers areas such as electrical insulation, electromagnetic compatibility, and protection against electric shock.
<b>ASTM Standards</b>	The American Society for Testing and Materials (ASTM) has developed a range of standards related to medical devices, covering areas such as materials, testing, and performance.
<b>ASTM E2224</b>	This standard outlines a procedure for evaluating the effectiveness of cleaning agents used in medical device manufacturing.
<b>GMP</b> (Good Manufacturing Practices)	This refers to the regulatory requirements for ensuring that products are consistently produced and controlled according to quality standards.
<b>USP &lt;788&gt;</b>	This is a general chapter in the United States Pharmacopeia that provides guidelines for particulate matter in injections and other pharmaceuticals.
<b>FDA QSR</b>	The FDA's Quality System Regulation requires medical device manufacturers to establish and maintain procedures for cleaning and sterilizing devices, including procedures for parts cleaning and rinsing.
<b>AAMI TIR12</b>	This technical information report from the Association for the Advancement of Medical Instrumentation provides guidance on the use of ultrasonic cleaners in medical device manufacturing.

These standards and regulations are designed to ensure that medical devices are safe, effective, and reliable for use by healthcare providers and patients. While particle counting is not specifically required by ISO 13485, it is a common practice in the industry and may be used as part of a broader quality management system to ensure that medical devices are produced in a clean and sterile environment.

**Validation of Cleaning Procedures:** The cleaning and disinfection procedures used in aseptic manufacturing must be validated to ensure that they are effective at removing dirt and killing microorganisms. This involves testing the effectiveness of the cleaning

and disinfection agents and procedures.

**Validation of Sterilization Procedures:** The sterilization procedures used to sterilize equipment must be validated to ensure that they are effective at killing microorganisms. This involves testing the equipment after sterilization to ensure that it is free from contamination.

**Validation of Personnel Procedures:** Personnel involved in aseptic manufacturing must follow strict hygiene procedures to minimize the risk of contamination. These procedures must be validated to ensure that they are effective.

# Conclusion

The manufacturing process for pharmaceutical injectable aseptic injections is complex and requires careful control and monitoring at every step. The process must be designed to ensure that the final product is sterile and free from contamination. Quality control testing is critical to ensure that the final product meets all regulatory requirements and is safe for patients to use.

Particle monitoring involves the use of specialized equipment, such as particle counters and microbial air samplers, to measure the number and size of particles in the air and on surfaces in the manufacturing environment. These measurements can be used to identify any potential sources of contamination and to take corrective actions to prevent contamination from occurring in the future.

# References

1. **"Guidance for Industry: Sterile Drug Products Produced by Aseptic Processing - Current Good Manufacturing Practice."** FDA, [www.fda.gov/regulatory-information/search-fda-guidance-documents/sterile-drug-products-produced-asepticprocessing-current-good-manufacturing-practice](http://www.fda.gov/regulatory-information/search-fda-guidance-documents/sterile-drug-products-produced-asepticprocessing-current-good-manufacturing-practice)
2. **"Aseptic Processing."** Parenteral Drug Association [www.pda.org/resources/technical-reports/aseptic-processing](http://www.pda.org/resources/technical-reports/aseptic-processing)
3. **"Sterilization Methods in the Pharmaceutical Industry."** American Pharmaceutical Review [www.americanpharmaceuticalreview.com/Featured-Articles/157417-Sterilization-Methods-in-the-Pharmaceutical-Industry](http://www.americanpharmaceuticalreview.com/Featured-Articles/157417-Sterilization-Methods-in-the-Pharmaceutical-Industry)
4. **"Validation of Sterilization Processes."** FDA [www.fda.gov/industry/study-validation/validation-sterilizationprocesses](http://www.fda.gov/industry/study-validation/validation-sterilizationprocesses)
5. **"Preparation of Personnel and Facilities for Aseptic Manufacturing."** Parenteral Drug Association [www.pda.org/docs/default-source/website-document-library/technical-reports/preparation-of-personnel-andfacilities-for-aseptic-manufacturing-1st-edition.pdf?sfvrsn=6a100d6f\\_4](http://www.pda.org/docs/default-source/website-document-library/technical-reports/preparation-of-personnel-andfacilities-for-aseptic-manufacturing-1st-edition.pdf?sfvrsn=6a100d6f_4)

Tel: 800.555.1212



[golighthouse.com](http://golighthouse.com)

©2025 Lighthouse Worldwide Solutions. Some rights reserved. Referenced products or services which are trademarked are owned by their respective trademark owners. No copyright infringement is intended, and is shared here only for educational purposes. Lighthouse Worldwide Solutions is based in White City, Oregon, USA with offices and service locations worldwide.